

# **Fanellan Hub 400 kV Substation and Converter Station**

## **Environmental Impact Assessment Report**

### **Volume 4 | Technical Appendices**





#### **Appendix 14.8 – Calibration Certificate NL74 34178103**

**February 2025**



## TECHNICAL APPENDIX 14: NOISE IMPACT ASSESSMENT

### 14.8 Calibration Certificate NC74 34178103

	<b>CERTIFICATE OF CALIBRATION</b>		 0653
<b>Date of Issue: 24 May 2022</b> Calibrated at & Certificate issued by: ANV Measurement Systems Beaufort Court 17 Roebuck Way Milton Keynes MK5 8HL Telephone 01908 642846 Fax 01908 642814 E-Mail: info@noise-and-vibration.co.uk Web: www.noise-and-vibration.co.uk <small>Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems</small>		<b>Certificate Number: UCRT22/1682</b>	
		Page 1 of 2 Pages	
		Approved Signatory  K. Mistry	
Customer	Wood Group St. Vincent Plaza (Floor 2) 319 St. Vincent Street Glasgow G2 5LP		
Order No.	26010406		
Test Procedure	Procedure TP 1 Calibration of Sound Calibrators		
Description	Acoustic Calibrator		
Identification	<i>Manufacturer</i> Rion	<i>Instrument</i> Calibrator	<i>Model</i> NC-74
			<i>Serial No.</i> 34178103
<p>The calibrator has been tested as specified in Annex B of IEC 60942:2003. As public evidence was available from a testing organisation (PTB) responsible for approving the results of pattern evaluation tests, to demonstrate that the model of sound calibrator fully conformed to the requirements for pattern evaluation described in Annex A of IEC 60942:2003, the sound calibrator tested is considered to conform to all the class 1 requirements of IEC 60942:2003.</p>			
ANV Job No.	UKAS22/05346		
Date Received	23 May 2022		
Date Calibrated	24 May 2022		
Previous Certificate	<i>Dated</i> 22 May 2020 <i>Certificate No.</i> UCRT20/1440 <i>Laboratory</i> 0653		
<p>This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.</p>			

## CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0653

Certificate Number

UCRT22/1682

Page 2 of 2 Pages

### Measurements

The sound pressure level generated by the calibrator in its WS2 configuration was measured five times by the Insert Voltage Method using a microphone as detailed below. The mean of the results obtained is shown below. It is corrected to the standard atmospheric pressure of 101.3 kPa (1013 mBar) using original manufacturers information.

Test Microphone	Manufacturer	Type
	Brüel & Kjær	4134

### Results

The level of the calibrator output under the conditions outlined above was

94.02 ± 0.10 dB rel 20 µPa

### Functional Tests and Observations

The frequency of the sound produced was	1001.97 ± 0.12 Hz
The total distortion was	1.61 ± 0.11 % Distortion

During the measurements environmental conditions were

Temperature	23 to 24 °C
Relative Humidity	44 to 51 %
Barometric Pressure	99.4 to 99.5 kPa

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

The uncertainties refer to the measured values only with no account being taken of the ability of the instrument to maintain its calibration.

A small correction factor may need to be applied to the sound pressure level quoted above if the device is used to calibrate a sound level meter which is fitted with a free-field response microphone. See manufacturers handbook for details.

END

### Note:

Calibrator adjusted prior to calibration?	NO
Initial Level	N/A dB
Initial Frequency	N/A Hz

Additional Comments The results on this certificate only relate to the items calibrated as identified above.  
None

Calibrated by: B. Bogdan

R 2